

10.0 510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

This summary of safety and effectiveness information is submitted in accordance with the requirements of 21 CFR 807.92.


Contact Person:	James P. Raskob LUNAR Corporation 726 Heartland Trail Madison, WI 53717
Phone:	(608) 826-7425
Fax:	(608) 826-7107
Date:	June 13 th , 2000
Device/Trade Name:	Pediatric Reference Data
Common Name:	Bone Densitometer
Classification Name:	Bone Densitometer 21CFR 892.1170
Predicate Devices:	Expanded LUNAR reference Data 510(k) submission K964307 NHANES Reference Data for QDR Bone Densitometers 510(k) submission K963363

10.1 DESCRIPTION OF THE DEVICE:

The Pediatric Reference Data expands the range of bone densitometry reference values to include ages 5 through 19 years of age. The software provides a comparison of measured variables obtained by dual energy x-ray absorptiometry to a database of reference values.

10.2 CONCLUSION

The Pediatric Reference Data is substantially equivalent to currently marketed software. No new safety and effectiveness questions are raised with the Pediatric Reference Data.



Signed

James P. Raskob

Printed Name

Regulatory Affairs/Quality Assurance Manager

Title



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

JUL 10 2000

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. James P. Raskob
Regulatory Affairs/Quality Assurance Manager
Lunar Corporation
726 Heartland Trail
Madison, WI 53717

Re: K001812
Pediatric Reference Data for Prodigy
and DPX Bone Densitometers
Dated: June 13, 2000
Received: June 15, 2000
Regulatory Class: II
21 CFR 892.1170/Procode: 90 KGI

Dear Mr. Raskob:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4591. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

Daniel G. Schultz, M.D.
Captain, USPHS
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure(s)

3.0 INDICATION FOR USE FORM

Page 1 of 1

- 510(k) Number (if known) K001812
- Device Name: Pediatric Reference Data
- Indications for use:

The Pediatric Reference Data is a software option used with the PRODIGY and DPX bone densitometers. This software expands the range of bone densitometry reference data to include ages 5 through 19 years of age. The software provides a comparison of measured variables obtained by dual energy x-ray absorptiometry to a database of reference values. These data can be used for comparative purposes at the sole discretion of the physician.

The use of the PRODIGY and DPX Bone Densitometers is restricted to prescription use only. The operator's manual for the PRODIGY and DPX contains the following statement:

"United States Federal Law restricts this device to the sale, distribution, and use by or on the order of a physician."

PLEASE DO NOT WRITE BELOW THIS LINE. CONTINUE ON ANOTHER PAGE IF NEEDED.

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓
(Per 21 CFR 801.109)

OR

Over-the-Counter Use _____
(Optional Format 1-2-96)

David A. Stymore
(Division Sign-Off)
Division of Reproductive, Abdominal, ENT,
and Radiological Devices
510(k) Number K001812